

Manual Supplement

Manual Title:	71X Series Calibration	Supplement Issue:	1
Part Number:	686540	Issue Date:	2/20
Print Date:	January 1998	Page Count:	1
Revision/Date:	12, 9/19		

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1, 34

On page 18, in Table 19, under **718 Version 4.0 or Higher, 718 Ex V 2.0 and Higher:**

Change the 4th column heading

From: **Accuracy** ^[1] **Pressure accuracy is \pm % of positive range**

To: **Accuracy** ^[1] **Pressure Accuracy is \pm % of Range**

On page 20, in Table 27, under the **Accuracy** ^[1] column:

Change

From: ± 0.025 % of positive Range (6 months calibration)

± 0.035 % of positive Range (1-year calibration)

To: ± 0.025 % of Range (6 months calibration)

± 0.035 % of Range (1-year calibration)