

## Instrument Security Procedures

**Model:** 718

**Product Name:** 718 300G, Pressure Calibrator 300 PSIG

**Instrument Description:** 4-20 mA loop calibrator, Pressure calibrator

**Memory Description:**

Microprocessor: MSP430F449

60KB+256B Flash Memory for calibration data and program memory

Microprocessor RAM:

2KB RAM

No Data logging or any measurement data storage.

**Memory Cleaning Instructions:**

Use Programming tool to erase microprocessor flash.