

Manual Supplement

Manual Title:	729/729 FC Calibration	Supplement Issue:	2
Part Number:	Web-only	Issue Date:	7/19
Print Date:	July 2017	Page Count:	1
Revision/Date:			

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1, 634

On page 6, under **Product Models**, add the following to the **Comment** column:

Dry air and non-corrosive gas only.

Change #2, 725

On page 12, replace the table with:

Input Pressure (psi)	12 Month Upper Limit	12 Month Lower Limit
729 300G/729 FC 300G		
-12	-11.94	-12.06
-9	-8.94	-9.06
-6	-5.94	-6.06
-3	-2.94	-3.06
0	0.06	-0.06
60	60.06	59.94
120	120.06	119.94
180	180.06	179.94
240	240.06	239.94
300	300.06	299.94
240	240.06	239.94
180	180.06	179.94
120	120.06	119.94
60	60.06	59.94
0	0.06	-0.06
-3	-2.94	-3.06
-6	-5.94	-6.06
-9	-8.94	-9.06
-12	-11.94	-12.06