

Manual Supplement

| | | | |
|----------------|----------------------------|-------------------|----------|
| Manual Title: | 123B/124B/125B Calibration | Supplement Issue: | 3 |
| Part Number: | Web-Only | Issue Date: | 9/17 |
| Print Date: | May 2016 | Page Count: | 1 |
| Revision/Date: | | | |

This supplement contains information necessary to ensure the accuracy of the above manual..

Change #1, 347

On page 40 , Table 17, change the **Input Value** for Cap Gain (CL 0566):

From: 50 nF

To: 500 nF

Change #2, 387, 613

On page 22, replace Step 6c with:

1. Noise filter: **OFF**

On page 23, replace Step 6g with:

2. Noise filter: **OFF**