

Statement of Memory Volatility

Product and Manufacturer				
Models: 1773/1775/1777 3 Phase Power Quality Analyzer		Manufacturer: Fluke Corporation		
		Address: 6920 Seaway Blvd.		
		City: Everett	State: WA	Zip: 98203
Volatile Memory				
Does the item contain volatile memory (i.e., memory whose contents are lost when power is removed)? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No (If the answer is 'Yes', provide the following information for each type of memory.)				
Memory Type (SRAM, DRAM, etc.):	Size:	User Modifiable:	Function:	Process to Clear:
Mobile DDR DRAM	2GB	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	Operating system, program execution and data storage	Power off device for at least 60 s
Processor caches	32kB-2MB	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	Operating system, program execution and data storage	Power off device for at least 60 s
STM32L010 NV memory	2kB	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	Power management program	Power off device for at least 60 s
USB-C controller int. SRAM		<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	USB-C controller firmware	Power off device for at least 60 s
FPGA RAM	504kbit	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	Measurement logic	Power off device for at least 60 s

Non-Volatile Memory

Does the item contain non-volatile memory (i.e., memory whose contents are retained when power is removed)?

Yes No (If the answer is 'Yes', provide the following information for each type of memory.)

Memory Type (BBRAM, Flash, EEPROM, etc.):	Size:	User Modifiable:	Function:	Process to Clear:
eMMC Flash	8GB	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	Operating system, program execution and data storage	Restore to Factory Defaults (see <i>Additional Information</i>)
Serial I2C EEPROM	1kB	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	Calibration data	N/A
TPM	6.8kB	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	Encryption keys	N/A
STM32L010 internal flash	16kB	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	Power management program	N/A
Flash for USB-C controller	8Mbit	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	USB-C controller application firmware	N/A
Flash for FPGA	256Mbit	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	FPGA firmware	N/A

Media

Does the item contain media storage capability (i.e., removable or non-removable disk drives, memory cards, etc.)?




Yes No (If the answer is 'Yes', provide the following information for each type of memory.)

	Storage Type (Disk, Tape, etc.):	Size:	User Modifiable:	Function:	Process to Clear:

Data storage extension (installed on 1777)	microSD card	32GB	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	Measurement data	Restore to Factory Defaults (see <i>Additional Information</i>)
---	--------------	------	---	------------------	---

Additional Information

To delete user data, such as measurement data, screenshots, and instrument settings:

1. Simultaneously push and hold    while the device starts until the **First-Time-Use** wizard displays.
2. Alternatively, use the *Reset to Factory Defaults* procedure described in the User Manual.
If a microSD card is installed, remove the microSD card from the device and do the *Reset to Factory Defaults* procedure.

Fluke Representative Information

Name: Frank Healy	Title: Product Manager	Office Phone: 425-446-5388	Fax/Email: frank.healy@fluke.com
-----------------------------	----------------------------------	--------------------------------------	--