

Instrument Security Procedures

Model:

Fluke Norma 4000/5000

Product Name:

Fluke Norma 4000/5000

Item number: EA1430Z / EA1560Z

Instrument Description:

Power Measurement Device

Memory Description:

Volatile: SDRAM MICRON MT48LC2M32B2P-6:G

Non-Volatile: Flash FUJITSU MBM29LV008BA-PTN
3x GAL LATTICE GAL22LV10C-7LJ

Memory Cleaning Instructions:

Volatile: will be erased by turning off the device

Non Volatile: open instrument and destroy component or destroy digital board
ZEA 1111 3501 10